

DART SENSORS LTD

DART MARINE PARK, TOTNES, DEVON TQ9 5AL ENGLAND

Tel/Fax: +44 (0)1803 868698

email: customer.service@dart-sensors.com

www.dart-sensors.com

Product Data Sheet: ethanol in water liquid standards

1. IDENTIFICATION OF SUBSTANCES

Ethanol in water solutions for breath simulators. Ethanol supplied by Merck Ltd. (*specification below*).

2. COMPOSITION DATA

35 microgram ethanol in water solution.

Accuracy required: 87.95 to 90.71 mg/100 cm³

Water	99.66 %
Ethanol	0.09%
CuSO ₄ .5H ₂ O	0.25% min.

50 microgram ethanol in water solution.

Accuracy required: 125.6 to 129.6 mg/100 cm³

Water	99.62 %
Ethanol	0.13%
CuSO ₄ .5H ₂ O	0.25% min.

140 microgram ethanol in water solution.

Accuracy required: 351.6 to 362.8 mg/100 cm³

Water	99.39%
Ethanol	0.36%
CuSO ₄ .5H ₂ O	0.25% min.

200 microgram ethanol in water solution.

Accuracy required: 502.6 to 518.3 mg/100 cm³

Water	99.24 %
Ethanol	0.51%
CuSO ₄ .5H ₂ O	0.25% min.

3. HANDLING & STORAGE

Avoid extremes of temperature. Store in a cool dark place.

Ethanol 99.7–100% v/v ("absolute") 'AnalaR'

C ₂ H ₅ OH	M.W.46.07
Minimum assay	99.8 (w/w) %
Identity	By IR spectrum
<i>Maximum Limits of Impurities</i>	
Colour	10 Hazen units
Water	0.2%
Acidity (CH ₃ COOH)	0.0006%
Alkalinity (NH ₃)	0.0002%
Non-volatile matter	0.001%
Acetone [(CH ₃) ₂ CO]	0.001%
Aldehydes and ketones (CH ₃ CHO)	0.001%
iso-Amyl alcohol [(CH ₃) ₂ CH(CH ₂) ₂ OH]	0.05%
Furfuraldehyde (C ₅ H ₄ O ₂)	0.001%
Fusel oil	Passes test
Higher alcohols (standard mixture)	0.02%
Methanol (CH ₃ OH)	0.01%
Propan-2-ol [(CH ₃) ₂ CHOH]	0.003%
Aluminium (Al)	0.00001%
Barium (Ba)	0.000002%
Cadmium (Cd)	0.000005%
Calcium (Ca)	0.00005%
Chromium (Cr)	0.000002%
Cobalt (Co)	0.000002%
Copper (Cu)	0.000002%
Iron (Fe)	0.00001%
Lead (Pb)	0.000005%
Magnesium (Mg)	0.000005%
Manganese (Mn)	0.000002%
Nickel (Ni)	0.000002%
Potassium (K)	0.00001%
Sodium (Na)	0.00005%
Strontium (Sr)	0.000002%
Tin (Sn)	0.00001%
Zinc (Zn)	0.00001%
Substances darkened by sulphuric acid	Passes test
Substances reducing permanganate (O)	0.0003%

Also complies with BS 6376: Part 2: 1984 (ISO 6353/2-1983) and with the corresponding ACS specification



1994

2004

Directors: W.J. King BSc., PhD.; S.M. King BSc.
Registered office: 30 Fore Street, Totnes TQ9 5RP
Registered in England & Wales no. 2931435
VAT Reg No 648 3625 15

1/1|1.1|04/05



Smart
Funding Innovation